

## CERTIFICATE OF ANALYSIS

PRODUCT NAME: Safe Connect 52°C

EVENT READING: 52°C

START MELT: 50.4°C

END MELT  
(Spec +/-1°C) 52.4°C

MID MELT: 51.4°C

BATCH NO: 54A19

CALIBRATION EQUIPMENT: Calibrated Heated Stage

Tested by: M. Lawford

Result: Pass